

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/10/2025

Lab File ID: BM049876.D Time Analyzed: 09:32

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	319220	7.781	1042650	10.57	653251	14.42
UPPER LIMIT	638440	8.281	2085300	11.074	1306500	14.921
LOWER LIMIT	159610	7.281	521325	10.074	326626	13.921
EPA SAMPLE NO.						
01 PB167509BL	304918	7.78	1031960	10.57	649646	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.