

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/09/2025

Lab File ID: BP024228.D Time Analyzed: 10:08

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	293576	7.769	1173450	10.55	712048	14.40
UPPER LIMIT	587152	8.269	2346900	11.052	1424100	14.904
LOWER LIMIT	146788	7.269	586725	10.052	356024	13.904
EPA SAMPLE NO.						
01 B-149-SB01	273538	7.74	1106820	10.52	675246	14.38
02 EB-2025-4-7	235704	7.74	911036	10.52	559100	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.