

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1746 SDG NO.: Q1746

Lab File ID: BM049875.D

DFTPP Injection Date: 04/10/2025

Instrument ID: BNA_M

DFTPP Injection Time: 08:53

| m/e | ION ABUNDANCE CRITERIA | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 51 | 10.0 - 80.0% of mass 198 | 21.6 |
| 68 | Less than 2.0% of mass 69 | 0.4 (1.4) 1 |
| 69 | Mass 69 relative abundance | 25.7 |
| 70 | Less than 2.0% of mass 69 | 0.1 (0.5) 1 |
| 127 | 10.0 - 80.0% of mass 198 | 33.2 |
| 197 | Less than 2.0% of mass 198 | 0.2 |
| 198 | Base Peak, 100% relative abundance | 100 |
| 199 | 5.0 to 9.0% of mass 198 | 6.8 |
| 275 | 10.0 - 60.0% of mass 198 | 25.9 |
| 365 | Greater than 1% of mass 198 | 3.4 |
| 441 | Present, but less than mass 443 | 11.7 |
| 442 | Greater than 50% of mass 198 | 100 |
| 443 | 15.0 - 24.0% of mass 442 | 15 (19.7) 2 |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA SAMPLE NO. | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| SSTDCCC040 | SSTDCCC040 | BM049876.D | 04/10/2025 | 09:32 |
| PB167509BL | PB167509BL | BM049877.D | 04/10/2025 | 10:11 |