

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1746 SDG NO.: Q1746

Lab File ID: BP024159.D

DFTPP Injection Date: 03/12/2025

Instrument ID: BNA_P

DFTPP Injection Time: 15:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.5
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	33.9
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.5
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	14.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024160.D	03/12/2025	16:17
SSTDICC005	SSTDICC005	BP024161.D	03/12/2025	16:57
SSTDICC010	SSTDICC010	BP024162.D	03/12/2025	17:38
SSTDICC020	SSTDICC020	BP024163.D	03/12/2025	18:19
SSTDICCC040	SSTDICCC040	BP024164.D	03/12/2025	19:00
SSTDICC050	SSTDICC050	BP024165.D	03/12/2025	19:41
SSTDICC060	SSTDICC060	BP024166.D	03/12/2025	20:21
SSTDICC080	SSTDICC080	BP024167.D	03/12/2025	21:02