

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bp041025

SequenceID : bp041025

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria	_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria	_____	_____✓_____	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____✓_____	_____
Compounds #32,77 and #85 are biased high in the CCC. If any samples are found with Hit of this compound they will be analyzed again with proper passing CCC.	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction	_____	_____✓_____	_____
d. Acid Fraction	_____	_____✓_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

\_\_\_\_ ✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

✓ \_\_\_\_\_

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

Internal standard failed in Q1748-08 which is not associated for required compounds.

\_\_\_\_\_ ✓ \_\_\_\_\_

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_\_ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_\_ ✓

ADDITIONAL COMMENTS:

anahy  
Analyst

04/10/2025  
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bp041025

SequenceID : bp041025 NA NO YES