

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP031225

Review By	anahy	Review On	3/13/2025 11:41:15 AM
Supervise By	Jagrut	Supervise On	3/17/2025 4:44:19 PM
SubDirectory	BP031225	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024159.D	12 Mar 2025 15:36		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024160.D	12 Mar 2025 16:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024161.D	12 Mar 2025 16:57	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024162.D	12 Mar 2025 17:38		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP024163.D	12 Mar 2025 18:19		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP024164.D	12 Mar 2025 19:00	This calibration is good for 8270E and 8270E DOD methods	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024165.D	12 Mar 2025 19:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024166.D	12 Mar 2025 20:21	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024167.D	12 Mar 2025 21:02	Compound #32,54,69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP031225	BP024168.D	12 Mar 2025 21:43		RC/JU	Ok,M
11	PB167078BL	PB167078BL	BP024169.D	12 Mar 2025 22:24		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041025

Review By	anahy	Review On	4/11/2025 10:19:20 AM		
Supervise By	Jagrut	Supervise On	4/11/2025 10:44:46 AM		
SubDirectory	BP041025	HP Acquire Method	BNA_P	HP Processing Method	bp031225
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12658,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024244.D	10 Apr 2025 09:05		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024245.D	10 Apr 2025 09:46		RC/JU	Ok
3	PB167518BL	PB167518BL	BP024246.D	10 Apr 2025 10:27		RC/JU	Ok
4	PB167537TB	PB167537TB	BP024247.D	10 Apr 2025 11:08		RC/JU	Ok
5	Q1745-04	IB-6A-WC	BP024248.D	10 Apr 2025 11:54		RC/JU	Ok
6	Q1745-12	IB-6.5-WC	BP024249.D	10 Apr 2025 12:35		RC/JU	Ok
7	Q1740-04	TP-20	BP024250.D	10 Apr 2025 13:15		RC/JU	Ok
8	Q1743-04	TP-16	BP024251.D	10 Apr 2025 13:56		RC/JU	Ok
9	Q1746-05	B-158-GW01	BP024252.D	10 Apr 2025 14:36		RC/JU	Ok
10	Q1746-06	B-149-GW01	BP024253.D	10 Apr 2025 15:17		RC/JU	Ok,M
11	Q1748-04	IB-1.5-WC	BP024254.D	10 Apr 2025 17:08		RC/JU	Ok
12	Q1748-08	IB-2A-WC	BP024255.D	10 Apr 2025 17:49		RC/JU	Ok
13	Q1754-02	TP-1	BP024256.D	10 Apr 2025 18:30		RC/JU	Ok
14	Q1754-04	TP-1-CONCRETE	BP024257.D	10 Apr 2025 19:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041125

Review By	Rahul	Review On	4/14/2025 11:10:32 AM
Supervise By	Jagrut	Supervise On	4/14/2025 3:53:13 PM
SubDirectory	BP041125	HP Acquire Method	BNA_P
		HP Processing Method	bp031225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12658,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024258.D	11 Apr 2025 09:02		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024259.D	11 Apr 2025 09:43	CCC fail high side for com. #32 and #77	RC/JU	Ok,M
3	PB167537BL	PB167537BL	BP024260.D	11 Apr 2025 10:23		RC/JU	Ok
4	PB167537BS	PB167537BS	BP024261.D	11 Apr 2025 11:04		RC/JU	Ok
5	PB167517TB	PB167517TB	BP024262.D	11 Apr 2025 11:45		RC/JU	Ok
6	PB167518BS	PB167518BS	BP024263.D	11 Apr 2025 12:25		RC/JU	Ok
7	Q1749-04	TP-14	BP024264.D	11 Apr 2025 13:38		RC/JU	Ok
8	Q1748-04MS	IB-1.5-WCMS	BP024265.D	11 Apr 2025 14:19		RC/JU	Ok,M
9	Q1748-04MSD	IB-1.5-WCMSD	BP024266.D	11 Apr 2025 15:00		RC/JU	Ok,M
10	Q1746-02	B-149-SB01	BP024267.D	11 Apr 2025 15:40		RC/JU	Ok
11	Q1746-04	B-149-SB02	BP024268.D	11 Apr 2025 16:21		RC/JU	Ok
12	Q1744-02	B-158-SB01	BP024269.D	11 Apr 2025 17:02		RC/JU	Ok
13	Q1744-04	B-158-SB02	BP024270.D	11 Apr 2025 17:42		RC/JU	Ok
14	Q1752-02	TP-1	BP024271.D	11 Apr 2025 18:23		RC/JU	Ok
15	Q1752-04	TP-2	BP024272.D	11 Apr 2025 19:04		RC/JU	Ok
16	Q1752-06	TP-4	BP024273.D	11 Apr 2025 19:45		RC/JU	Ok

M : Manual Integration