

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1746

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

DataFile: BP024261.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167537BS	Pyridine	50	42.6	ug/L	85				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	49.4	ug/L	99				70 (76)	130 (103)	
	2-Methylphenol	50	54.4	ug/L	109				70 (69)	130 (109)	
	3+4-Methylphenols	50	53.8	ug/L	108				20 (67)	160 (106)	
	Hexachloroethane	50	50.3	ug/L	101				20 (76)	160 (118)	
	Nitrobenzene	50	51.1	ug/L	102				70 (58)	130 (106)	
	Hexachlorobutadiene	50	50.0	ug/L	100				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	53.7	ug/L	107				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	53.9	ug/L	108				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	59.4	ug/L	119				70 (60)	130 (115)	
	Hexachlorobenzene	50	50.4	ug/L	101				70 (73)	130 (106)	
	Pentachlorophenol	100	110	ug/L	110				20 (47)	160 (114)	