

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG No.: Q1746  
 Instrument ID: MSVOA\_X Calibration Date(s): 04/01/2025 04/01/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02  
 GC Column: DB-624UI ID: 0.18 (mm)

| LAB FILE ID:          |        | RRF001 = VX045525.D |        | RRF005 = VX045526.D |        | RRF020 = VX045527.D |       |       |  |
|-----------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                       |        | RRF050 = VX045528.D |        | RRF100 = VX045529.D |        | RRF150 = VX045530.D |       |       |  |
| COMPOUND              | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| Vinyl Chloride        | 0.671  | 0.662               | 0.701  | 0.716               | 0.738  | 0.730               | 0.703 | 4.4   |  |
| 1,1-Dichloroethene    | 0.563  | 0.588               | 0.612  | 0.604               | 0.618  | 0.616               | 0.600 | 3.5   |  |
| 2-Butanone            | 0.474  | 0.537               | 0.582  | 0.586               | 0.571  | 0.548               | 0.550 | 7.5   |  |
| Carbon Tetrachloride  | 0.450  | 0.497               | 0.521  | 0.536               | 0.545  | 0.556               | 0.518 | 7.5   |  |
| Chloroform            | 1.244  | 1.309               | 1.348  | 1.315               | 1.309  | 1.309               | 1.306 | 2.6   |  |
| Benzene               | 1.414  | 1.416               | 1.519  | 1.481               | 1.483  | 1.465               | 1.463 | 2.8   |  |
| 1,2-Dichloroethane    | 0.533  | 0.585               | 0.649  | 0.622               | 0.620  | 0.617               | 0.604 | 6.7   |  |
| Trichloroethene       | 0.349  | 0.322               | 0.356  | 0.351               | 0.351  | 0.356               | 0.348 | 3.7   |  |
| Tetrachloroethene     | 0.346  | 0.373               | 0.371  | 0.347               | 0.333  | 0.347               | 0.353 | 4.5   |  |
| Chlorobenzene         | 0.951  | 1.054               | 1.123  | 1.084               | 1.086  | 1.112               | 1.068 | 5.8   |  |
| 1,2-Dichloroethane-d4 |        | 0.962               | 0.900  | 0.868               | 0.904  | 0.937               | 0.914 | 3.9   |  |
| Dibromofluoromethane  |        | 0.372               | 0.342  | 0.345               | 0.353  | 0.362               | 0.355 | 3.5   |  |
| Toluene-d8            |        | 1.257               | 1.233  | 1.214               | 1.230  | 1.257               | 1.238 | 1.5   |  |
| 4-Bromofluorobenzene  |        | 0.413               | 0.448  | 0.453               | 0.481  | 0.460               | 0.451 | 5.5   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.