

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG No.: Q1746

Instrument ID: MSVOA_Y Calibration Date/Time: 04/10/2025 09:51

Lab File ID: VY021826.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:08 12:02

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.473		-8.69	20
Chloromethane	0.729	0.722	0.1	-0.96	20
Vinyl Chloride	0.824	0.777		-5.7	20
Bromomethane	0.562	0.510		-9.25	20
Chloroethane	0.539	0.515		-4.45	20
Trichlorofluoromethane	1.056	0.995		-5.78	20
1,1,2-Trichlorotrifluoroethane	0.557	0.548		-1.62	20
1,1-Dichloroethene	0.524	0.521		-0.57	20
Acetone	0.111	0.105		-5.41	20
Carbon Disulfide	1.727	1.642		-4.92	20
Methyl tert-butyl Ether	1.306	1.226		-6.13	20
Methyl Acetate	0.275	0.298		8.36	20
Methylene Chloride	0.626	0.577		-7.83	20
trans-1,2-Dichloroethene	0.577	0.585		1.39	20
1,1-Dichloroethane	1.049	1.074	0.1	2.38	20
Cyclohexane	0.956	0.885		-7.43	20
2-Butanone	0.158	0.140		-11.39	20
Carbon Tetrachloride	0.566	0.596		5.3	20
cis-1,2-Dichloroethene	0.649	0.671		3.39	20
Bromochloromethane	0.442	0.448		1.36	20
Chloroform	1.091	1.115		2.2	20
1,1,1-Trichloroethane	0.986	1.001		1.52	20
Methylcyclohexane	0.590	0.615		4.24	20
Benzene	1.479	1.589		7.44	20
1,2-Dichloroethane	0.417	0.422		1.2	20
Trichloroethene	0.377	0.407		7.96	20
1,2-Dichloropropane	0.350	0.377		7.71	20
Bromodichloromethane	0.523	0.551		5.35	20
4-Methyl-2-Pentanone	0.230	0.211		-8.26	20
Toluene	0.923	1.022		10.73	20
t-1,3-Dichloropropene	0.465	0.471		1.29	20
cis-1,3-Dichloropropene	0.548	0.564		2.92	20
1,1,2-Trichloroethane	0.260	0.265		1.92	20
2-Hexanone	0.153	0.145		-5.23	20
Dibromochloromethane	0.355	0.370		4.22	20
1,2-Dibromoethane	0.244	0.244		0	20
Tetrachloroethene	0.455	0.482		5.93	20
Chlorobenzene	1.146	1.216	0.3	6.11	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG No.: Q1746

Instrument ID: MSVOA_Y Calibration Date/Time: 04/10/2025 09:51

Lab File ID: VY021826.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:08 12:02

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.965	2.166		10.23	20
m/p-Xylenes	0.753	0.846		12.35	20
o-Xylene	0.695	0.776		11.65	20
Styrene	1.158	1.305		12.69	20
Bromoform	0.232	0.227	0.1	-2.15	20
Isopropylbenzene	3.714	3.986		7.32	20
1,1,2,2-Tetrachloroethane	0.668	0.605	0.3	-9.43	20
1,3-Dichlorobenzene	1.757	1.868		6.32	20
1,4-Dichlorobenzene	1.736	1.812		4.38	20
1,2-Dichlorobenzene	1.529	1.605		4.97	20
1,2-Dibromo-3-Chloropropane	0.103	0.091		-11.65	20
1,2,4-Trichlorobenzene	0.829	0.913		10.13	20
1,2,3-Trichlorobenzene	0.707	0.777		9.9	20
1,2-Dichloroethane-d4	0.542	0.533		-1.66	20
Dibromofluoromethane	0.324	0.356		9.88	20
Toluene-d8	1.234	1.384		12.16	20
4-Bromofluorobenzene	0.417	0.462		10.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.