

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040725

Review By	John Carlone	Review On	4/8/2025 10:03:01 AM
Supervise By	Maresh Dadoda	Supervise On	4/8/2025 2:05:33 PM
SubDirectory	VX040725	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133600
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133601,VP133602

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045613.D	07 Apr 2025 08:34		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045614.D	07 Apr 2025 09:17		JC/MD	Ok,M
3	VX0407MBL01	VX0407MBL01	VX045615.D	07 Apr 2025 09:47	pH#Lot#V12668	JC/MD	Ok
4	VX0407WBL01	VX0407WBL01	VX045616.D	07 Apr 2025 10:11		JC/MD	Ok
5	VX0407WBS01	VX0407WBS01	VX045617.D	07 Apr 2025 10:34		JC/MD	Ok,M
6	VX0407WBSD01	VX0407WBSD01	VX045618.D	07 Apr 2025 11:00		JC/MD	Ok,M
7	Q1731-04	RMW-03B-90-040325	VX045619.D	07 Apr 2025 11:23	vial A pH<2	JC/MD	Ok
8	Q1731-01	RMW-01B-82-040325	VX045620.D	07 Apr 2025 11:46	vial A pH<2	JC/MD	Ok,M
9	Q1731-02	RMW-04B-91-040325	VX045621.D	07 Apr 2025 12:09	vial A pH<2	JC/MD	Ok
10	Q1731-03	RMW-01B-82-040325-F	VX045622.D	07 Apr 2025 12:33	vial A pH<2	JC/MD	Ok
11	Q1741-03	FB01-040425	VX045623.D	07 Apr 2025 12:56	vial A pH<2,FB	JC/MD	Ok
12	Q1741-04	TB01-040425	VX045624.D	07 Apr 2025 13:19	vial A pH<2,TB	JC/MD	Ok
13	Q1741-01	RMW-02B-66-040425	VX045625.D	07 Apr 2025 13:43	run 50x	JC/MD	Not Ok
14	Q1741-02	RMW-06B-74-040425	VX045626.D	07 Apr 2025 14:06	Need Straight Run	JC/MD	Not Ok
15	IBLK	IBLK	VX045627.D	07 Apr 2025 14:35		JC/MD	Ok
16	Q1741-02	RMW-06B-74-040425	VX045628.D	07 Apr 2025 14:58	vial A pH<2	JC/MD	Ok
17	Q1746-08	TB-2025-4-7	VX045629.D	07 Apr 2025 15:21	vial A pH<2 TB	JC/MD	Ok
18	Q1746-07	EB-2025-4-7	VX045630.D	07 Apr 2025 15:45	vial A pH<2 EB,Hit of com.#16	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040725

Review By	John Carlone	Review On	4/8/2025 10:03:01 AM
Supervise By	Maresh Dadoda	Supervise On	4/8/2025 2:05:33 PM
SubDirectory	VX040725	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133600
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133601,VP133602

19	Q1746-05	B-158-GW01	VX045631.D	07 Apr 2025 16:08	vial A pH<2	JC/MD	Ok
20	Q1746-06	B-149-GW01	VX045632.D	07 Apr 2025 16:31	vial A pH<2	JC/MD	Ok
21	VSTDCCC050	VSTDCCC050EC	VX045633.D	07 Apr 2025 16:55		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040825

Review By	John Carlone	Review On	4/9/2025 9:21:19 AM
Supervise By	Maresh Dadoda	Supervise On	4/9/2025 11:41:10 AM
SubDirectory	VX040825	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133619
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133620,VP133621

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045634.D	08 Apr 2025 07:49		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045635.D	08 Apr 2025 08:24	pH#Lot#V12668	JC/MD	Ok,M
3	VX0408MBL01	VX0408MBL01	VX045636.D	08 Apr 2025 08:49		JC/MD	Ok
4	VX0408WBL01	VX0408BWL01	VX045637.D	08 Apr 2025 09:12		JC/MD	Ok
5	VX0408WBS01	VX0408WBS01	VX045638.D	08 Apr 2025 09:43		JC/MD	Ok,M
6	VX0408WBSD01	VX0408WBSD01	VX045639.D	08 Apr 2025 10:08		JC/MD	Ok,M
7	Q1741-01	RMW-02B-66-040425	VX045640.D	08 Apr 2025 10:32	vial A pH<2	JC/MD	Ok
8	Q1746-07	EB-2025-4-7	VX045641.D	08 Apr 2025 10:55	vial B pH<2 EB	JC/MD	Ok
9	Q1702-03	60442	VX045642.D	08 Apr 2025 11:18	vial A pH<2 Foamy sample	JC/MD	Ok
10	PB167492ZHE01	PB167492ZHE01	VX045643.D	08 Apr 2025 11:41		JC/MD	Ok
11	PB167492ZHE02	PB167492ZHE02	VX045644.D	08 Apr 2025 12:05		JC/MD	Ok
12	Q1737-02	RT3069	VX045645.D	08 Apr 2025 12:28	vial A pH#5.0	JC/MD	Ok
13	PB167492ZHE04	PB167492ZHE04	VX045646.D	08 Apr 2025 12:51		JC/MD	Ok
14	PB167492ZHE05	PB167492ZHE05	VX045647.D	08 Apr 2025 13:14		JC/MD	Ok
15	PB167492ZHE06	PB167492ZHE06	VX045648.D	08 Apr 2025 13:38		JC/MD	Ok
16	PB167492ZHE07	PB167492ZHE07	VX045649.D	08 Apr 2025 14:01		JC/MD	Ok
17	PB167492ZHE08	PB167492ZHE08	VX045650.D	08 Apr 2025 14:24		JC/MD	Ok
18	PB167492ZHE09	PB167492ZHE09	VX045651.D	08 Apr 2025 14:47		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040825

Review By	John Carlone	Review On	4/9/2025 9:21:19 AM
Supervise By	Maresh Dadoda	Supervise On	4/9/2025 11:41:10 AM
SubDirectory	VX040825	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133619		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133620,VP133621		

19	PB167492ZHE10	PB167492ZHE10	VX045652.D	08 Apr 2025 15:11		JC/MD	Ok
20	PB167492ZHE11	PB167492ZHE11	VX045653.D	08 Apr 2025 15:34		JC/MD	Ok
21	PB167492ZHE03	PB167492ZHE03	VX045654.D	08 Apr 2025 15:58		JC/MD	Ok
22	Q1745-04	IB-6A-WC	VX045655.D	08 Apr 2025 16:21	vial A pH#5.0	JC/MD	Ok
23	Q1732-04	TT-8	VX045656.D	08 Apr 2025 16:44	vial A pH#5.0	JC/MD	Ok
24	Q1740-04	TP-20	VX045657.D	08 Apr 2025 17:07	vial A pH#5.0	JC/MD	Ok
25	Q1745-12	IB-6.5-WC	VX045658.D	08 Apr 2025 17:31	vial A pH#5.0	JC/MD	Ok
26	Q1743-04	TP-16	VX045659.D	08 Apr 2025 17:54	vial A pH#5.0	JC/MD	Ok
27	Q1744-02	B-158-SB01	VX045660.D	08 Apr 2025 18:17	vial A pH#5.0	JC/MD	Ok
28	Q1744-04	B-158-SB02	VX045661.D	08 Apr 2025 18:40	vial A pH#5.0	JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX045662.D	08 Apr 2025 19:04		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133499
Initial Calibration Stds	VP133500,VP133501,VP133502,VP133503,VP133505,VP133508
CCC	VP133510
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133509
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBS01	VY0327SBS01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041025

Review By	Maresh Dadoda	Review On	4/11/2025 1:59:08 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/11/2025 2:00:45 PM
SubDirectory	VY041025	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133628
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133629,VP133630 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021825.D	10 Apr 2025 08:34		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021826.D	10 Apr 2025 09:51		SY/MD	Ok,M
3	VY0410SBL01	VY0410SBL01	VY021827.D	10 Apr 2025 10:28		SY/MD	Ok
4	VY0410SBS01	VY0410SBS01	VY021828.D	10 Apr 2025 10:59		SY/MD	Ok,M
5	VY0410SBSD01	VY0410SBSD01	VY021829.D	10 Apr 2025 11:22		SY/MD	Ok,M
6	Q1747-01RE	ARS20-0008RE	VY021830.D	10 Apr 2025 12:06	vial-B not purged	SY/MD	Not Ok
7	Q1754-01RE	TP-1RE	VY021831.D	10 Apr 2025 12:29	vial-B Internal standard fail;Not match for com.#64	SY/MD	Not Ok
8	Q1732-03	TT-8-VOC	VY021832.D	10 Apr 2025 12:53	vial-A	SY/MD	Ok
9	Q1746-01	B-149-SB01	VY021833.D	10 Apr 2025 13:16	vial-A	SY/MD	Ok
10	Q1746-03	B-149-SB02	VY021834.D	10 Apr 2025 13:40	vial-A	SY/MD	Ok
11	Q1744-01	B-158-SB01	VY021835.D	10 Apr 2025 14:03	vial-A	SY/MD	Ok
12	Q1744-03	B-158-SB02	VY021836.D	10 Apr 2025 14:26	vial-A	SY/MD	Ok
13	Q1738-01RE	OK-02-040425RE	VY021837.D	10 Apr 2025 14:50	vial-B Internal Standard Fail	SY/MD	Confirms
14	Q1749-03	TP-14-VOC	VY021838.D	10 Apr 2025 15:13	vial-A	SY/MD	Ok
15	Q1743-03	TP-16-VOC	VY021839.D	10 Apr 2025 15:37	vial-A	SY/MD	Ok
16	Q1752-01	TP-1	VY021840.D	10 Apr 2025 16:00	vial-A	SY/MD	Ok
17	Q1752-03	TP-2	VY021841.D	10 Apr 2025 16:24	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041025

Review By	Mahesh Dadoda	Review On	4/11/2025 1:59:08 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/11/2025 2:00:45 PM		
SubDirectory	VY041025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133628				
CCC	VP133629,VP133630				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1752-05	TP-4	VY021842.D	10 Apr 2025 16:47	vial-A	SY/MD	Ok
19	Q1752-07	TP-3	VY021843.D	10 Apr 2025 17:10	vial-A	SY/MD	Ok
20	Q1771-03	TP-6C-VOC	VY021844.D	10 Apr 2025 17:34	vial-A	SY/MD	Ok
21	Q1771-07	TP-7C-VOC	VY021845.D	10 Apr 2025 17:57	vial-A	SY/MD	Ok
22	Q1771-11	TP-3A-VOC	VY021846.D	10 Apr 2025 18:21	vial-A	SY/MD	Ok
23	Q1760-03	TP-17-VOC	VY021847.D	10 Apr 2025 18:44	vial-A	SY/MD	Ok
24	Q1760-07	TP-15-VOC	VY021848.D	10 Apr 2025 19:08	vial-A Internal Standard Fail; Surrogate Fail;Need MeOH	SY/MD	Dilution

M : Manual Integration