

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746
Lab File ID: VX045635.D Date Analyzed: 04/08/2025
Instrument ID: MSVOA_X Time Analyzed: 08:24
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	102556	5.54	175513	6.75	152677	10.05
UPPER LIMIT	205112	6.038	351026	7.251	305354	10.549
LOWER LIMIT	51278	5.038	87756.5	6.251	76338.5	9.549
EPA SAMPLE NO.						
EB-2025-4-7	65792	5.54	130088	6.76	121595	10.05
VX0408WBL01	66539	5.55	129800	6.76	123174	10.05
VX0408WBS01	106087	5.54	183473	6.75	162323	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.