

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746
 Lab File ID: VY021826.D Date Analyzed: 04/10/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	147243	13.353				
UPPER LIMIT	294486	13.853				
LOWER LIMIT	73621.5	12.853				
EPA SAMPLE NO.						
B-149-SB01	79360	13.35				
B-149-SB02	109046	13.35				
VY0410SBL01	131635	13.35				
VY0410SBS01	144045	13.35				
VY0410SBSD01	148690	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.