

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: Q1746 SAS No.: Q1746 SDG NO.: Q1746
Lab File ID: VX045634.D BFB Injection Date: 04/08/2025
Instrument ID: MSVOA_X BFB Injection Time: 07:49
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.9
75	30.0 - 60.0% of mass 95	58.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.9
175	5.0 - 9.0% of mass 174	5.2 (7.7) 1
176	95.0 - 101.0% of mass 174	63.8 (95.3) 1
177	5.0 - 9.0% of mass 176	4.8 (7.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045635.D	04/08/2025	08:24
VX0408WBL01	VX0408WBL01	VX045637.D	04/08/2025	09:12
VX0408WBS01	VX0408WBS01	VX045638.D	04/08/2025	09:43
EB-2025-4-7	Q1746-07	VX045641.D	04/08/2025	10:55