

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc. **SDG No.:** Q1750
Contract: SAFE04 **Lab Code:** CHEM **Case No.:** Q1750 **SAS No.:** Q1750
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	973	1000	97	95 - 105	P	04/15/2025	15:47	1b135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc. **SDG No.:** Q1750
Contract: SAFE04 **Lab Code:** CHEM **Case No.:** Q1750 **SAS No.:** Q1750
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV	Lead	12.1	12.0	101	80 - 120	P	04/15/2025	15:51	1b135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc. **SDG No.:** Q1750
Contract: SAFE04 **Lab Code:** CHEM **Case No.:** Q1750 **SAS No.:** Q1750
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4940	5000	99	90 - 110	P	04/15/2025	16:26	lb135445
CCV02	Lead	4820	5000	96	90 - 110	P	04/15/2025	17:16	lb135445
CCV03	Lead	4910	5000	98	90 - 110	P	04/15/2025	18:07	lb135445
CCV04	Lead	4950	5000	99	90 - 110	P	04/15/2025	19:08	lb135445
CCV05	Lead	4830	5000	97	90 - 110	P	04/15/2025	19:59	lb135445
CCV06	Lead	4960	5000	99	90 - 110	P	04/15/2025	20:51	lb135445
CCV07	Lead	4950	5000	99	90 - 110	P	04/15/2025	21:43	lb135445
CCV08	Lead	4940	5000	99	90 - 110	P	04/15/2025	22:25	lb135445