



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q1750

Contract: SAFE04

Lab Code: CHEM

Case No.: Q1750

SAS No.: Q1750

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI1	Lead	12.5	12.0	104	40 - 160	P	04/15/2025	16:05	1b135445