

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Safety Environmental Co. of NY, Inc. **SDG No.:** Q1750
Contract: SAFE04 **Lab Code:** CHEM **Case No.:** Q1750 **SAS No.:** Q1750
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	11.2			-12	12	04/15/2025	16:09	lb135445
ICSAB01	Lead	41.0	49.0	84	37	61	04/15/2025	16:13	lb135445
ICSA	Lead	3.69			-12	12	04/15/2025	16:18	lb135445
ICSAB	Lead	49.2	49.0	100	37	61	04/15/2025	16:22	lb135445