

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Safety Environmental Co. of NY, Inc. **level:** low **sdg no.:** Q1750
contract: SAFE04 **lab code:** CHEM **case no.:** Q1750 **sas no.:** Q1750
matrix: Water **sample id:** Q1750-08 **client id:** CLASSROOM-4-2ND-RUNMS
Percent Solids for Sample: NA **Spiked ID:** Q1750-08MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	394		6.00	U	500	79		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Safety Environmental Co. of NY, Inc. **level:** low **sdg no.:** Q1750
contract: SAFE04 **lab code:** CHEM **case no.:** Q1750 **sas no.:** Q1750
matrix: Water **sample id:** Q1750-08 **client id:** CLASSROOM-4-2ND-RUNMSD
Percent Solids for Sample: NA **Spiked ID:** Q1750-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Lead	ug/L	75 - 125	374		6.00	U	500	75	N	P