

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1757

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042289.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0409ABS02	Vinyl Chloride	10	8.70	ppbv	87			70	130	
	Heptane	10	9.70	ppbv	97			70	130	
	1,1-Dichloroethene	10	8.90	ppbv	89			70	130	
	Cyclohexane	10	9.40	ppbv	94			70	130	
	cis-1,2-Dichloroethene	10	9.70	ppbv	97			70	130	
	1,1,1-Trichloroethane	10	8.90	ppbv	89			70	130	
	2,2,4-Trimethylpentane	10	10.7	ppbv	107			70	130	
	Benzene	10	10.6	ppbv	106			70	130	
	Trichloroethene	10	10.1	ppbv	101			70	130	
	Toluene	10	9.90	ppbv	99			70	130	
	Tetrachloroethene	10	8.90	ppbv	89			70	130	
	Ethyl Benzene	10	10.6	ppbv	106			70	130	
	m/p-Xylene	20	22.2	ppbv	111			70	130	
	o-Xylene	10	11.0	ppbv	110			70	130	
	1,3,5-Trimethylbenzene	10	10.9	ppbv	109			70	130	
	1,2,4-Trimethylbenzene	10	10.7	ppbv	107			70	130	
	Naphthalene	10	11.3	ppbv	113			70	130	
	Hexane	10	9.90	ppbv	99			70	130	