

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135364

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP112659
Chlorine Calibration std, 0.1ppm	WP112654
Chlorine Calibration std, 0.2ppm	WP112655
Chlorine Calibration std, 0.8ppm	WP112656
Chlorine Calibration std, 0.0ppm	WP112653
Chlorine Calibration std, 1.6ppm	WP112657
Residual Chlorine Calibration and CCV std, 0	WP112658
Total Chlorine Powder Pillows	W3147

Intercept: 0.0020

Slope: 0.9413

Regression: 0.999672

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		04/10/2025	08:50
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	4	04/10/2025	08:53
3	CAL3	0.2	1	0.000	0.190	0.190	0.20	0	04/10/2025	08:56
4	CAL4	0.4	1	0.000	0.360	0.360	0.38	-5	04/10/2025	08:59
5	CAL5	0.8	1	0.000	0.780	0.780	0.83	3.4	04/10/2025	09:02
6	CAL6	1.6	1	0.000	1.500	1.500	1.59	-0.6	04/10/2025	09:05

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135364

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3800	0.3800	0.3960	04/10/2025	09:08
2	ICB				1	0.0000	0.0000	0.0000	-0.0090	04/10/2025	09:11
3	CCV1			0.4	1	0.0000	0.3700	0.3700	0.3910	04/10/2025	09:14
4	CCB1				1	0.0000	0.0000	0.0000	-0.0090	04/10/2025	09:17
5	LB135364BL	50	50		1	0.0000	0.0000	0.0000	-0.0090	04/10/2025	09:20
6	LB135364BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.4070	04/10/2025	09:23
7	Q1759-01	50	50		1	0.0000	0.2500	0.2500	0.2580	04/10/2025	09:26
8	Q1759-01DUP	50	50		1	0.0000	0.2500	0.2500	0.2580	04/10/2025	09:29
9	Q1759-01MS	50	50	0.4	1	0.0000	0.6300	0.6300	0.6620	04/10/2025	09:32
10	Q1759-01MSD	50	50	0.4	1	0.0000	0.6200	0.6200	0.6510	04/10/2025	09:35
11	Q1759-02	50	50		1	0.0000	0.1400	0.1400	0.1410	04/10/2025	09:38
12	Q1759-03	50	50		1	0.0000	0.2100	0.2100	0.2150	04/10/2025	09:41
13	Q1759-04	50	50		1	0.0000	0.0300	0.0300	0.0230	04/10/2025	09:44
14	Q1759-05	50	50		1	0.0000	0.0800	0.0800	0.0770	04/10/2025	09:47
15	CCV2			0.4	1	0.0000	0.3600	0.3600	0.3800	04/10/2025	09:50
16	CCB2				1	0.0000	0.0000	0.0000	-0.0090	04/10/2025	09:53
17	Q1759-06	50	50		1	0.0000	0.1500	0.1500	0.1510	04/10/2025	09:56
18	Q1759-07	50	50		1	0.0000	0.2800	0.2800	0.2900	04/10/2025	09:59
19	CCV3			0.4	1	0.0000	0.3500	0.3500	0.3700	04/10/2025	10:02
20	CCB3				1	0.0000	0.0000	0.0000	-0.0090	04/10/2025	10:05

LB135364

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RESID CHLOR-0410202		WorkList ID : 188841		Department : Wet-Chemistry		Date : 04-10-2025 08:19:13		
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1759-01 - B	B2-W8-RM1016-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-02 - B	B2-W8-RM1010-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-03 - B	B2-W8-RM1011-RHSINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-04 - B	B2-W8-LH-DRINKING-FOUNTA	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-05 - B	B2-W8-RM1019-EYEWASH	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-06 - B	B2-W8-RM1040-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G
Q1759-07 - B	B2-W8-RM1041-SINK	Water	Residual Chlorine	Cool 4 deg C	CLOU03	L41	04/09/2025	SM4500 CI G

Date/Time 04.10.2025, 08:25
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: [Signature]

Date/Time 04.10.2025, 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: NF(wc)