



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: CHA Companies, Inc.

SDG No.: Q1759

Project: NYSOMH Emergency Service

RunNo.: LB135364

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	04/10/2025
Sample ID: CCB1 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	04/10/2025
Sample ID: CCB2 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	04/10/2025
Sample ID: CCB3 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.023	0.1	04/10/2025

Initial and Continuing Calibration Blank Summary

Client: CHA Companies, Inc.

SDG No.: Q1759

Project: NYSOMH Emergency Service

RunNo.: LB135374

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Turbidity	NTU	0.282	0.5000	J	0.15	1.0	04/10/2025
Sample ID: CCB1 Turbidity	NTU	0.276	0.5000	J	0.15	1	04/10/2025
Sample ID: CCB2 Turbidity	NTU	0.276	0.5000	J	0.15	1	04/10/2025