

Matrix Spike Summary

Client:	CHA Companies, Inc.	SDG No.:	Q1759
Project:	NYSOMH Emergency Service	Sample ID:	Q1759-01
Client ID:	B2-W8-RM1016-SINKMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.66		0.26		0.4	1	101		04/10/2025

Matrix Spike Summary

Client:	CHA Companies, Inc.	SDG No.:	Q1759
Project:	NYSOMH Emergency Service	Sample ID:	Q1759-01
Client ID:	B2-W8-RM1016-SINKMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Residual Chlorine	mg/L	71-148	0.65		0.26		0.4	1	98		04/10/2025