

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133499
Initial Calibration Stds	VP133500,VP133501,VP133502,VP133503,VP133505,VP133508
CCC	VP133510
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133509
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBS01	VY0327SBS01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041125

Review By	Maresh Dadoda	Review On	4/14/2025 3:42:48 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/14/2025 3:45:15 PM
SubDirectory	VY041125	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133637		
CCC	VP133638,VP133639		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021849.D	11 Apr 2025 08:46		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021850.D	11 Apr 2025 09:18		SY/MD	Ok,M
3	VY0411SBL01	VY0411SBL01	VY021851.D	11 Apr 2025 11:35		SY/MD	Ok
4	VY0411SBS01	VY0411SBS01	VY021852.D	11 Apr 2025 12:06		SY/MD	Ok,M
5	VY0411SBSD01	VY0411SBSD01	VY021853.D	11 Apr 2025 12:28		SY/MD	Ok,M
6	Q1756-03	TP-9-VOC	VY021854.D	11 Apr 2025 13:10	vial-A Internal Standard Fail ;Surrogate fail	SY/MD	ReRun
7	Q1761-01	GST3	VY021855.D	11 Apr 2025 13:33	vial-A	SY/MD	Ok
8	Q1781-03	WC-13-VOC	VY021856.D	11 Apr 2025 13:57	vial-A	SY/MD	Ok
9	Q1781-07	WC-12-VOC	VY021857.D	11 Apr 2025 14:20	vial-A	SY/MD	Ok
10	Q1781-11	WC-11-VOC	VY021858.D	11 Apr 2025 14:44	vial-A	SY/MD	Ok
11	Q1768-01	CRUSHED STONE	VY021859.D	11 Apr 2025 15:07	vial-A Internal Standard Fail ;Surrogate fail	SY/MD	ReRun
12	Q1779-07	TP-23-VOC	VY021860.D	11 Apr 2025 15:31	vial-A	SY/MD	Ok
13	Q1783-01	SU-03-41125	VY021861.D	11 Apr 2025 15:54	vial-A Internal Standard Fail	SY/MD	ReRun
14	Q1784-01	OR-01-41125	VY021862.D	11 Apr 2025 16:18	vial-A Internal Standard Fail	SY/MD	ReRun
15	VSTDCCC050	VSTDCCC050EC	VY021863.D	11 Apr 2025 16:40		SY/MD	Ok,M

M : Manual Integration