

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1761 SAS No.: Q1761 SDG NO.: Q1761
 Lab File ID: VY021850.D Date Analyzed: 04/11/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	196263	13.346				
UPPER LIMIT	392526	13.846				
LOWER LIMIT	98131.5	12.846				
EPA SAMPLE NO.						
GST3	180059	13.35				
VY0411SBL01	174550	13.35				
VY0411SBS01	190311	13.35				
VY0411SBSD01	188335	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.