

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX040925

SequenceID : VX040925

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	<u>✓</u>	<u>      </u>	<u>      </u>
2. GC/MS Tuning Specifications. BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	<u>✓</u>	<u>      </u>	<u>      </u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	<u>✓</u>	<u>      </u>	<u>      </u>
4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	<u>✓</u>	<u>      </u>	<u>      </u>
5. GC/MS Calibration Requirements	<u>✓</u>	<u>      </u>	<u>      </u>
5a. Initial Calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u>      </u>	<u>      </u>
5b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u>      </u>	<u>      </u>
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	<u>✓</u>	<u>      </u>	<u>      </u>
7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u>      </u>	<u>      </u>
8a. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	<u>      </u>	<u>      </u>
8b. Blank Spike/Blank spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	<u>      </u>	<u>      </u>
9. Internal Standard Area/Retention Time Shift Meet Criteria	<u>✓</u>	<u>      </u>	<u>      </u>
Comments:			

JOHN

Analyst

04/09/2025

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX040925

SequenceID : VX040925

NA NO YES

10. Analysis Holding Time Met

✓              

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

Q1739-01,02. These samples were run at straight dilution due to samples having oil layer in vial. Analyst checked with supervisor for dilution factor.

JOHN

Analyst

04/09/2025

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX040925

SequenceID : VX040925 NA NO YES

JOHN  
Analyst

04/09/2025  
Date