

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01  
Lab Code: CHEM Case No.: Q1762 SAS No.: Q1762 SDG NO.: Q1762  
Lab File ID: VX045664.D Date Analyzed: 04/09/2025  
Instrument ID: MSVOA\_X Time Analyzed: 10:35  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 98594         | 5.54  | 170453        | 6.75  | 147839        | 10.05  |
| UPPER LIMIT    | 197188        | 6.044 | 340906        | 7.251 | 295678        | 10.549 |
| LOWER LIMIT    | 49297         | 5.044 | 85226.5       | 6.251 | 73919.5       | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MW4            | 73129         | 5.55  | 135470        | 6.76  | 125359        | 10.05  |
| MW5            | 66105         | 5.54  | 129673        | 6.76  | 121124        | 10.05  |
| VX0409WBL01    | 69921         | 5.54  | 138334        | 6.76  | 127856        | 10.05  |
| VX0409WBS01    | 94074         | 5.54  | 163361        | 6.76  | 141398        | 10.05  |
| VX0409WBSD01   | 83922         | 5.54  | 147335        | 6.76  | 129399        | 10.05  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.