

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1762 SAS No.: Q1762 SDG NO.: Q1762
 Lab File ID: VX045712.D Date Analyzed: 04/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 02:49
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83221	5.54	146715	6.75	129670	10.05
UPPER LIMIT	166442	6.037	293430	7.251	259340	10.549
LOWER LIMIT	41610.5	5.037	73357.5	6.251	64835	9.549
EPA SAMPLE NO.						
MW4DL	63356	5.55	122653	6.76	112188	10.06
VX0410WBL02	68793	5.54	134259	6.76	125713	10.05
VX0410WBS02	87130	5.54	156738	6.76	139704	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.