

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1762 SAS No.: Q1762 SDG NO.: Q1762
 Lab File ID: VX045712.D Date Analyzed: 04/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 02:49
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63235	12.018				
UPPER LIMIT	126470	12.518				
LOWER LIMIT	31617.5	11.518				
EPA SAMPLE NO.						
MW4DL	47681	12.02				
VX0410WBL02	55453	12.02				
VX0410WBS02	64250	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.