

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1762 SAS No.: Q1762 SDG NO.: Q1762
Lab File ID: VX045663.D BFB Injection Date: 04/09/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:38
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.2
75	30.0 - 60.0% of mass 95	59.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1 (1.4) 1
174	50.0 - 100.0% of mass 95	67.4
175	5.0 - 9.0% of mass 174	5.1 (7.5) 1
176	95.0 - 101.0% of mass 174	66.1 (98) 1
177	5.0 - 9.0% of mass 176	4.4 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045664.D	04/09/2025	10:35
VX0409WBL01	VX0409WBL01	VX045666.D	04/09/2025	11:26
VX0409WBS01	VX0409WBS01	VX045667.D	04/09/2025	11:49
VX0409WBSD01	VX0409WBSD01	VX045668.D	04/09/2025	12:14
MW4	Q1762-01	VX045681.D	04/09/2025	17:16
MW5	Q1762-02	VX045682.D	04/09/2025	17:39