

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

RunNo.: LB135457

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	50.319	50	101	95-105	04/16/2025
Sample ID: CCV2 COD	mg/L	50.319	50	101	95-105	04/16/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

RunNo.: LB135462

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.01	7	100	90-110	04/10/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7	100	90-110	04/10/2025
Sample ID:	CCV2						
Field pH		pH	7.00	7	100	90-110	04/10/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

RunNo.: LB135462

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------