

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1766
Project:	Linden Generating Station - LRSA	Sample ID:	Q1766-01
Client ID:	DSN002MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	110		72.1		50.0	1	76		04/16/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1766
Project:	Linden Generating Station - LRSA	Sample ID:	Q1766-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	112		72.1		50.0	1	80		04/16/2025