

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

Run No.: LB135378

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135378BS							
BOD5	mg/L	198	191		97	1	84.6-115.4	04/10/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

Run No.: LB135387

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135387BS							
TSS	mg/L	550	532		97	1	90-110	04/11/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1766

Project: Linden Generating Station - LRSA

Run No.: LB135457

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135457BS							
COD	mg/L	50	47.4		95	1	90-110	04/16/2025