

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1769
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1769 **SAS No.:** Q1769
ICS Source: EPA **Instrument ID:** P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silver	0.030			-2	2	04/11/2025	13:10	LB135403
ICSAB01	Silver	18.3	18.0	102	15.3	20.7	04/11/2025	13:28	LB135403
ICSA01	Silver	0.040			-2	2	05/13/2025	16:19	LB135757
ICSAB1	Silver	18.1	18.0	101	15.3	20.7	05/13/2025	16:29	LB135757