

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1770

Project: Monthly 2025

Run No.: LB135378

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135378BS							
BOD5	mg/L	198	191		97	1	84.6-115.4	04/10/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1770

Project: Monthly 2025

Run No.: LB135387

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135387BS							
TSS	mg/L	550	532		97	1	90-110	04/11/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1770

Project: Monthly 2025

Run No.: LB135466

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135466BS							
TPH	mg/L	20.0	16.7		84	1	78-114	04/17/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1770

Project: Monthly 2025

Run No.: LB135466

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135466BSD							
TPH	mg/L	20.0	16.9		84	1	78-114	04/17/2025