

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1772 SAS No.: Q1772 SDG No.: Q1772

Instrument ID: MSVOA_L Calibration Date/Time: 04/10/2025 09:27

Lab File ID: VL042300.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.639	0.538		-15.81	30
1,1-Dichloroethene	0.552	0.476		-13.77	30
cis-1,2-Dichloroethene	1.423	1.294		-9.06	30
1,1,1-Trichloroethane	2.349	2.102		-10.52	30
Trichloroethene	0.414	0.436		5.31	30
Tetrachloroethene	0.393	0.363		-7.63	30
1-Bromo-4-Fluorobenzene	0.821	0.814		-0.85	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.