

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1773 SAS No.: Q1773 SDG No.: Q1773

Instrument ID: MSVOA_X Calibration Date/Time: 04/11/2025 10:41

Lab File ID: VX045730.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
t-1,3-Dichloropropene	50.000	47.58	4.832001	20