

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1773 SAS No.: Q1773 SDG NO.: Q1773
Lab File ID: VX045730.D Date Analyzed: 04/11/2025
Instrument ID: MSVOA_X Time Analyzed: 10:41
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97283	5.54	164714	6.76	147741	10.05
UPPER LIMIT	194566	6.044	329428	7.257	295482	10.549
LOWER LIMIT	48641.5	5.044	82357	6.257	73870.5	9.549
EPA SAMPLE NO.						
BP-TB-20250407	64635	5.54	127456	6.76	119779	10.05
BP-TT190D2-GW-20250407	64518	5.55	124304	6.76	112481	10.05
BP-TT190D1-GW-20250409	67136	5.55	131053	6.76	122019	10.05
BP-TT190D1-GW-20250409DL	66514	5.54	129658	6.76	118906	10.05
VX0411WBL01	65554	5.55	128675	6.76	118153	10.05
VX0411WBS01	92540	5.54	163903	6.76	140888	10.05
VX0411WBSD01	84320	5.54	150272	6.76	131860	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.