

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135391

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/11/2025	09:15
2	CAL2	1	Water	NA	NA	20.2	7.00	04/11/2025	09:16
3	CAL3	1	Water	NA	NA	20.3	10.02	04/11/2025	09:17
4	ICV	1	Water	NA	NA	20.3	7.01	04/11/2025	09:20
5	CCV1	1	Water	NA	NA	20.3	2.01	04/11/2025	09:25
6	Q1763-02	1	Water	NA	NA	20.4	5.66	04/11/2025	09:33
7	Q1763-02DUP	1	Water	NA	NA	20.5	5.67	04/11/2025	09:34
8	Q1764-02	1	Water	NA	NA	20.2	6.17	04/11/2025	09:37
9	Q1767-01	1	Water	NA	NA	20.3	6.87	04/11/2025	09:44
10	Q1774-02	1	Water	NA	NA	20.2	8.73	04/11/2025	10:00
11	Q1774-03	1	Water	NA	NA	20.3	7.05	04/11/2025	10:10
12	Q1774-04	1	Water	NA	NA	20.9	6.62	04/11/2025	10:15
13	Q1782-01	1	Water	NA	NA	20.1	5.95	04/11/2025	10:30
14	Q1782-03	1	Water	NA	NA	20.2	6.82	04/11/2025	10:35
15	Q1782-05	1	Water	NA	NA	20.3	6.71	04/11/2025	10:40
16	CCV2	1	Water	NA	NA	20.2	12.02	04/11/2025	10:41
17	Q1782-07	1	Water	NA	NA	20.1	6.68	04/11/2025	10:45
18	CCV3	1	Water	NA	NA	20.3	2.01	04/11/2025	10:47

WORKLIST(Hardcopy Internal Chain)

JB 135391

WorkList Name : ph w q1764

WorkList ID : 188859

Department : Wet-Chemistry

Date : 04-11-2025 07:52:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1763-02	AUD-25-0047	Water	pH	Cool 4 deg C	PSEG03	I31	04/10/2025	9040C
Q1764-02	BUR-25-COMP	Water	pH	Cool 4 deg C	PSEG03	L41	04/10/2025	9040C
Q1767-01	NWB-2152	Water	pH	Cool 4 deg C	PSEG03	F11	04/10/2025	9040C
Q1774-02	TT-073-IDWGW-20250409	Water	pH	Cool 4 deg C	TETR06	F11	04/09/2025	9040C
Q1774-03	TT-074-IDWGW-20250409	Water	pH	Cool 4 deg C	TETR06	F11	04/09/2025	9040C
Q1774-04	TT-075-IDWGW-20250409	Water	pH	Cool 4 deg C	TETR06	F11	04/09/2025	9040C
Q1782-01	MW-1	Water	pH	Cool 4 deg C	LOCK01	K11	04/09/2025	9040C
Q1782-03	MW-2	Water	pH	Cool 4 deg C	LOCK01	K11	04/09/2025	9040C
Q1782-05	MW-3	Water	pH	Cool 4 deg C	LOCK01	K11	04/09/2025	9040C
Q1782-07	MW-4	Water	pH	Cool 4 deg C	LOCK01	K11	04/09/2025	9040C

Date/Time 04-11-25 09:10

Raw Sample Received by: JB wcc

Raw Sample Relinquished by: CR SM

Date/Time 04-11-25

Raw Sample Received by: CR SM

Raw Sample Relinquished by: JB wcc