

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q1774

Project: CTO WE13

RunNo.: LB135391

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	04/11/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	04/11/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	04/11/2025
Sample ID: CCV3 pH	pH	2.01	2.00	101	90-110	04/11/2025