

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	04/09/25
Project:	CTO WE13	Date Received:	04/10/25
Client Sample ID:	TT-075-IDWGW-20250409	SDG No.:	Q1774
Lab Sample ID:	Q1774-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	16600		1	28.3	40.0	50.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-36-0	Antimony	6.25	U	1	2.06	6.25	25.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-38-2	Arsenic	7.46	J	1	3.48	8.00	10.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-39-3	Barium	102	N	1	6.28	12.5	50.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-41-7	Beryllium	1.65	J	1	0.13	0.75	3.00	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-43-9	Cadmium	0.75	U	1	0.094	0.75	3.00	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-70-2	Calcium	12900		1	33.0	250	1000	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-47-3	Chromium	38.6	N	1	0.66	2.50	5.00	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-48-4	Cobalt	9.44	J	1	0.50	3.75	15.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-50-8	Copper	30.6	N	1	7.07	8.00	10.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7439-89-6	Iron	14100		1	18.5	40.0	50.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7439-92-1	Lead	29.9		1	3.51	4.80	6.00	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7439-95-4	Magnesium	4550	N	1	39.4	250	1000	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7439-96-5	Manganese	518	N	1	1.46	2.50	10.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7439-97-6	Mercury	0.16	UN	1	0.076	0.16	0.20	ug/L	04/10/25 14:20	04/11/25 11:19	SW7470A	
7440-02-0	Nickel	22.8		1	0.85	5.00	20.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-09-7	Potassium	2280		1	685	800	1000	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7782-49-2	Selenium	8.00	U	1	5.88	8.00	10.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-22-4	Silver	2.50	U	1	0.58	2.50	5.00	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-23-5	Sodium	25000		1	237	500	1000	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-28-0	Thallium	10.0	U	1	2.32	10.0	20.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-62-2	Vanadium	19.5	JN	1	3.06	10.0	20.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010
7440-66-6	Zinc	49.3	N	1	1.75	5.00	20.0	ug/L	04/11/25 10:15	04/14/25 15:14	SW6010	SW3010

Color Before:	light Brown	Clarity Before:	Cloudy	Texture:
Color After:	light Brown	Clarity After:	Cloudy	Artifacts:
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits