

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1774
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1774 **SAS No.:** Q1774

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167567BS							
Aluminum	ug/L	1000	948		95	86 - 115	P
Antimony	ug/L	400	405		101	88 - 113	P
Arsenic	ug/L	400	401		100	87 - 113	P
Barium	ug/L	100	96.6		97	88 - 113	P
Beryllium	ug/L	100	95.2		95	89 - 112	P
Cadmium	ug/L	100	101		101	88 - 113	P
Calcium	ug/L	500	504	J	101	87 - 113	P
Chromium	ug/L	200	199		100	90 - 113	P
Cobalt	ug/L	100	98.3		98	89 - 114	P
Copper	ug/L	150	154		103	86 - 114	P
Iron	ug/L	1500	1530		102	87 - 115	P
Lead	ug/L	500	500		100	86 - 113	P
Magnesium	ug/L	1000	943	J	94	85 - 113	P
Manganese	ug/L	100	99.1		99	90 - 114	P
Nickel	ug/L	250	246		98	88 - 113	P
Potassium	ug/L	5000	4890		98	86 - 114	P
Selenium	ug/L	1000	1010		101	83 - 114	P
Silver	ug/L	37.5	36.6		98	84 - 115	P
Sodium	ug/L	1500	1430		95	87 - 115	P
Thallium	ug/L	1000	1030		103	85 - 114	P
Vanadium	ug/L	150	145		97	90 - 111	P
Zinc	ug/L	100	99.1		99	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1774

Contract: TETR06

Lab Code: CHEM

Case No.: Q1774

SAS No.: Q1774

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167569BS Mercury	ug/L	4.0	4.04		101	82 - 119	CV