

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

Continuing Calib Date: 04/10/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 09:33 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.67	5.66	5.56	5.76	-0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.14	6.04	6.24	0.00
Aroclor-1260-1	(1)	7.26	7.26	7.16	7.36	0.00
Aroclor-1260-2	(2)	7.51	7.51	7.41	7.61	0.00
Aroclor-1260-3	(3)	7.87	7.87	7.77	7.97	0.00
Aroclor-1260-4	(4)	8.10	8.09	7.99	8.19	-0.01
Aroclor-1260-5	(5)	8.42	8.41	8.31	8.51	-0.01
Tetrachloro-m-xylene		4.51	4.51	4.41	4.61	0.00
Decachlorobiphenyl		10.23	10.23	10.13	10.33	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

Continuing Calib Date: 04/10/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 09:33 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-3	(3)	5.10	5.10	5.00	5.20	0.01
Aroclor-1016-4	(4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5	(5)	5.35	5.36	5.26	5.46	0.01
Aroclor-1260-1	(1)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-2	(2)	6.57	6.58	6.48	6.68	0.01
Aroclor-1260-3	(3)	6.73	6.73	6.63	6.83	0.00
Aroclor-1260-4	(4)	7.20	7.21	7.11	7.31	0.01
Aroclor-1260-5	(5)	7.44	7.45	7.35	7.55	0.01
Tetrachloro-m-xylene		3.81	3.82	3.72	3.92	0.01
Decachlorobiphenyl		8.85	8.86	8.76	8.96	0.01