

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP071206.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.670	5.563	5.763	496.930	500.000	-0.6
Aroclor-1016-2	5.691	5.585	5.785	495.850	500.000	-0.8
Aroclor-1016-3	5.753	5.648	5.848	482.230	500.000	-3.6
Aroclor-1016-4	5.850	5.745	5.945	517.440	500.000	3.5
Aroclor-1016-5	6.144	6.038	6.238	526.010	500.000	5.2
Aroclor-1260-1	7.263	7.156	7.356	525.350	500.000	5.1
Aroclor-1260-2	7.516	7.410	7.610	529.730	500.000	5.9
Aroclor-1260-3	7.876	7.768	7.968	531.440	500.000	6.3
Aroclor-1260-4	8.099	7.992	8.192	470.160	500.000	-6.0
Aroclor-1260-5	8.419	8.311	8.511	492.520	500.000	-1.5
Decachlorobiphenyl	10.234	10.125	10.325	50.660	50.000	1.3
Tetrachloro-m-xylene	4.516	4.412	4.612	52.980	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 04/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP071206.D Time Analyzed: 16:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.803	5.003	546.170	500.000	9.2
Aroclor-1016-2	4.918	4.822	5.022	529.800	500.000	6.0
Aroclor-1016-3	5.094	4.999	5.199	562.150	500.000	12.4
Aroclor-1016-4	5.136	5.041	5.241	566.440	500.000	13.3
Aroclor-1016-5	5.351	5.255	5.455	568.390	500.000	13.7
Aroclor-1260-1	6.386	6.292	6.492	546.080	500.000	9.2
Aroclor-1260-2	6.574	6.480	6.680	523.580	500.000	4.7
Aroclor-1260-3	6.727	6.634	6.834	530.430	500.000	6.1
Aroclor-1260-4	7.198	7.105	7.305	502.190	500.000	0.4
Aroclor-1260-5	7.440	7.346	7.546	440.140	500.000	-12.0
Decachlorobiphenyl	8.850	8.759	8.959	43.700	50.000	-12.6
Tetrachloro-m-xylene	3.813	3.716	3.916	53.310	50.000	6.6