

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP071217.D Time Analyzed: 21:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.668	5.563	5.763	522.540	500.000	4.5
Aroclor-1016-2	5.689	5.585	5.785	521.640	500.000	4.3
Aroclor-1016-3	5.752	5.648	5.848	509.140	500.000	1.8
Aroclor-1016-4	5.849	5.745	5.945	557.900	500.000	11.6
Aroclor-1016-5	6.142	6.038	6.238	577.980	500.000	15.6
Aroclor-1260-1	7.259	7.156	7.356	565.900	500.000	13.2
Aroclor-1260-2	7.515	7.410	7.610	582.410	500.000	16.5
Aroclor-1260-3	7.873	7.768	7.968	549.310	500.000	9.9
Aroclor-1260-4	8.099	7.992	8.192	529.790	500.000	6.0
Aroclor-1260-5	8.417	8.311	8.511	541.830	500.000	8.4
Decachlorobiphenyl	10.233	10.125	10.325	53.220	50.000	6.4
Tetrachloro-m-xylene	4.514	4.412	4.612	53.300	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 04/10/2025

Lab Sample No.: AR1660CCC500 Data File : PP071217.D Time Analyzed: 21:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.803	5.003	535.360	500.000	7.1
Aroclor-1016-2	4.917	4.822	5.022	523.850	500.000	4.8
Aroclor-1016-3	5.094	4.999	5.199	561.950	500.000	12.4
Aroclor-1016-4	5.136	5.041	5.241	553.160	500.000	10.6
Aroclor-1016-5	5.350	5.255	5.455	577.870	500.000	15.6
Aroclor-1260-1	6.386	6.292	6.492	576.420	500.000	15.3
Aroclor-1260-2	6.574	6.480	6.680	552.720	500.000	10.5
Aroclor-1260-3	6.727	6.634	6.834	561.900	500.000	12.4
Aroclor-1260-4	7.198	7.105	7.305	515.220	500.000	3.0
Aroclor-1260-5	7.440	7.346	7.546	465.030	500.000	-7.0
Decachlorobiphenyl	8.851	8.759	8.959	44.970	50.000	-10.1
Tetrachloro-m-xylene	3.812	3.716	3.916	49.440	50.000	-1.1