

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/11/2025

Lab Sample No.: AR1660CCC500 Data File : PP071223.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.675	5.563	5.763	449.470	500.000	-10.1
Aroclor-1016-2	5.697	5.585	5.785	424.640	500.000	-15.1
Aroclor-1016-3	5.759	5.648	5.848	452.500	500.000	-9.5
Aroclor-1016-4	5.856	5.745	5.945	464.800	500.000	-7.0
Aroclor-1016-5	6.149	6.038	6.238	483.520	500.000	-3.3
Aroclor-1260-1	7.269	7.156	7.356	518.500	500.000	3.7
Aroclor-1260-2	7.523	7.410	7.610	509.860	500.000	2.0
Aroclor-1260-3	7.881	7.768	7.968	472.810	500.000	-5.4
Aroclor-1260-4	8.105	7.992	8.192	453.220	500.000	-9.4
Aroclor-1260-5	8.426	8.311	8.511	467.610	500.000	-6.5
Decachlorobiphenyl	10.245	10.125	10.325	57.330	50.000	14.7
Tetrachloro-m-xylene	4.522	4.412	4.612	45.120	50.000	-9.8

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 04/11/2025

Lab Sample No.: AR1660CCC500 Data File : PP071223.D Time Analyzed: 09:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.803	5.003	476.160	500.000	-4.8
Aroclor-1016-2	4.920	4.822	5.022	457.000	500.000	-8.6
Aroclor-1016-3	5.097	4.999	5.199	491.450	500.000	-1.7
Aroclor-1016-4	5.140	5.041	5.241	489.550	500.000	-2.1
Aroclor-1016-5	5.353	5.255	5.455	512.450	500.000	2.5
Aroclor-1260-1	6.388	6.292	6.492	505.110	500.000	1.0
Aroclor-1260-2	6.577	6.480	6.680	479.210	500.000	-4.2
Aroclor-1260-3	6.731	6.634	6.834	479.920	500.000	-4.0
Aroclor-1260-4	7.202	7.105	7.305	449.200	500.000	-10.2
Aroclor-1260-5	7.443	7.346	7.546	403.620	500.000	-19.3
Decachlorobiphenyl	8.855	8.759	8.959	48.360	50.000	-3.3
Tetrachloro-m-xylene	3.815	3.716	3.916	45.510	50.000	-9.0