

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN041125

Review By	John Carlone	Review On	4/14/2025 9:15:54 AM
Supervise By	Semsettin Yesilyurt	Supervise On	4/14/2025 3:45:12 PM
SubDirectory	VN041125	HP Acquire Method	HP Processing Method 624N041125W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133643		
Initial Calibration Stds	VP133645,VP133646,VP133647,VP133648,VP133649		
CCC	VP133644		
Internal Standard/PEM			
ICV/I.BLK	VP133650		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN086241.D	11 Apr 2025 09:52	JC\MD	Ok
2	VSTDICC100	VN086242.D	11 Apr 2025 11:04	JC\MD	Ok,M
3	VSTDICC050	VN086243.D	11 Apr 2025 11:28	JC\MD	Not Ok
4	VSTDICC050	VN086244.D	11 Apr 2025 11:52	JC\MD	Ok,M
5	VSTDICCC020	VN086245.D	11 Apr 2025 12:16	JC\MD	Ok,M
6	VSTDICC005	VN086246.D	11 Apr 2025 12:40	JC\MD	Ok,M
7	IBLK	VN086247.D	11 Apr 2025 13:04	JC\MD	Ok
8	VSTDICV020	VN086248.D	11 Apr 2025 13:34	JC\MD	Ok,M
9	VN0411WBS01	VN086249.D	11 Apr 2025 14:38	JC\MD	Ok,M
10	VN0411WBSD01	VN086250.D	11 Apr 2025 15:16	JC\MD	Ok,M
11	VN0411WBL01	VN086251.D	11 Apr 2025 15:40	JC\MD	Ok
12	Q1774-02	VN086252.D	11 Apr 2025 16:15	JC\MD	Ok
13	Q1774-03	VN086253.D	11 Apr 2025 16:39	JC\MD	Ok,M
14	Q1774-04	VN086254.D	11 Apr 2025 17:03	JC\MD	Ok,M
15	VSTDCCC020	VN086255.D	11 Apr 2025 17:28	JC\MD	Ok,M

M : Manual Integration