

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1774 SAS No.: Q1774 SDG NO.: Q1774
Lab File ID: VN086245.D Date Analyzed: 04/11/2025
Instrument ID: MSVOA_N Time Analyzed: 12:16
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	30974	7.81	185020	9.10	163907	11.87
UPPER LIMIT	61948	8.312	370040	9.6	327814	12.365
LOWER LIMIT	15487	7.312	92510	8.6	81953.5	11.365
EPA SAMPLE NO.						
TT-073-IDWGW-20250409	16665	7.82	98006	9.10	84075	11.87
TT-074-IDWGW-20250409	19774	7.82	120171	9.09	102960	11.87
TT-075-IDWGW-20250409	29637	7.81	180090	9.10	151244	11.87
VN0411WBL01	32111	7.81	197127	9.10	168421	11.87
VN0411WBS01	32103	7.81	201467	9.10	175632	11.87
VN0411WBSD01	34735	7.81	202525	9.10	180363	11.86

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.