

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1775

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135408

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID: CCV1 TOC	mg/L	973	1000	97	90-110	04/14/2025
Sample ID: CCV2 TOC	mg/L	1080	1000	108	90-110	04/14/2025
Sample ID: CCV3 TOC	mg/L	1020	1000	102	90-110	04/14/2025