

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1778

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135456

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	1030	1000	103	90-110	03/14/2025
Sample ID:	CCV1						
TOC		mg/L	965	1000	96	90-110	04/17/2025
Sample ID:	CCV2						
TOC		mg/L	1000	1000	100	90-110	04/17/2025
Sample ID:	CCV3						
TOC		mg/L	1040	1000	104	90-110	04/17/2025
Sample ID:	CCV4						
TOC		mg/L	1000	1000	100	90-110	04/17/2025