

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	517	510	102	95 - 105	P	04/10/2025	10:09	LB135373
	Iron	10300	10000	103	95 - 105	P	04/10/2025	10:09	LB135373
	Lead	951	1000	95	95 - 105	P	04/10/2025	10:09	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.9	20.0	110	80 - 120	P	04/10/2025	10:25	LB135373
	Iron	89.9	100	90	80 - 120	P	04/10/2025	10:25	LB135373
	Lead	9.94	12.0	83	80 - 120	P	04/10/2025	10:25	LB135373

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	101	90 - 110	P	04/10/2025	11:00	LB135373
	Iron	5250	5000	105	90 - 110	P	04/10/2025	11:00	LB135373
	Lead	5240	5000	105	90 - 110	P	04/10/2025	11:00	LB135373
CCV02	Copper	1270	1250	102	90 - 110	P	04/10/2025	12:08	LB135373
	Iron	5230	5000	104	90 - 110	P	04/10/2025	12:08	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	12:08	LB135373
CCV03	Copper	1290	1250	103	90 - 110	P	04/10/2025	12:59	LB135373
	Iron	5230	5000	105	90 - 110	P	04/10/2025	12:59	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	12:59	LB135373
CCV04	Copper	1290	1250	103	90 - 110	P	04/10/2025	13:51	LB135373
	Iron	4980	5000	100	90 - 110	P	04/10/2025	13:51	LB135373
	Lead	4760	5000	95	90 - 110	P	04/10/2025	13:51	LB135373
CCV05	Copper	1280	1250	103	90 - 110	P	04/10/2025	14:44	LB135373
	Iron	5160	5000	103	90 - 110	P	04/10/2025	14:44	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	14:44	LB135373
CCV06	Copper	1270	1250	102	90 - 110	P	04/10/2025	15:44	LB135373
	Iron	5060	5000	101	90 - 110	P	04/10/2025	15:44	LB135373
	Lead	4690	5000	94	90 - 110	P	04/10/2025	15:44	LB135373
CCV07	Copper	1290	1250	103	90 - 110	P	04/10/2025	16:56	LB135373
	Iron	4820	5000	96	90 - 110	P	04/10/2025	16:56	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	16:56	LB135373
CCV08	Copper	1280	1250	103	90 - 110	P	04/10/2025	17:56	LB135373
	Iron	5030	5000	100	90 - 110	P	04/10/2025	17:56	LB135373
	Lead	4740	5000	95	90 - 110	P	04/10/2025	17:56	LB135373
CCV09	Copper	1290	1250	103	90 - 110	P	04/10/2025	18:19	LB135373
	Iron	5130	5000	103	90 - 110	P	04/10/2025	18:19	LB135373
	Lead	4720	5000	94	90 - 110	P	04/10/2025	18:19	LB135373