



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.9	20.0	109	40 - 160	P	04/10/2025	10:33	LB135373
	Iron	102	100	102	40 - 160	P	04/10/2025	10:33	LB135373
	Lead	11.6	12.0	97	40 - 160	P	04/10/2025	10:33	LB135373